

MMET, Inc.
Middleton Microbiological & Environmental Testing Laboratory
 3889 N. 20th Street, Ozark, MO 65721
 417-581-MMET (6638), Toll Free 877-581-MMET, Fax 417-582-0269
 Email: mmetinc@classicnet.net

Laboratory Report

Report Number: M12119	Report Date: 3/5/2008
Lab Number: 80249	
<hr/>	
Customer: Additive Technologies, LLC 1515 W 2200 S., Suite C Salt Lake City, UT 84119	Project Manager: Jennifer Geib
Phone: 303-300-6525	Project Name: Cleaner Evaluation
Fax: 801-340-0217	Sample Matrix: Concentrated Cleaner
Cell:	Sampled By: Jennifer Geib
Email:	Sample ID: Clean Safe C3
Purchase Order No.:	Date Sampled: 2/24/2008
	Date Received: 2/25/2008 1053
<hr/>	
Analysis Requested: Volatiles by EPA Method 8260.	

Analyte	Result	Detection Limit	Units	Date of Analysis	Analyst
1, 1 - Dichloroethene	< 50	50.000	mg/L	2/29/2008	GCL
1, 2 - Dichloroethane	< 50	50.000	mg/L	2/29/2008	GCL
1, 4 - Dichlorobenzene	< 50	50.000	mg/L	2/29/2008	GCL
Benzene	< 50	50.000	mg/L	2/29/2008	GCL
Carbon tetrachloride	< 50	50.000	mg/L	2/29/2008	GCL
Chlorobenzene	< 50	50.000	mg/L	2/29/2008	GCL
Chloroform	< 50	50.000	mg/L	2/29/2008	GCL
MEK (2 - Butanone)	< 100	100.000	mg/L	2/29/2008	GCL
Tetrachloroethene	< 50	50.000	mg/L	2/29/2008	GCL
Trichloroethene	< 50	50.000	mg/L	2/29/2008	GCL
Vinyl chloride	< 100	100.000	mg/L	2/29/2008	GCL
Surrogate Recovery	Percent Recovery	Date of Analysis	Analyst		
Dibromochloromethane	97	2/29/2008	GCL		
1,2-Dichloroethane-d4	97	2/29/2008	GCL		
Toluene-d8	98	2/29/2008	GCL		
Bromofluorobenzene	89	2/29/2008	GCL		

Conclusion: No volatile compounds were detected at the dilution analyzed. It was necessary to perform a 10000x dilution to avoid foaming of the sample during analysis



1997-2007

Report Approved by: _____
 Wayne A. Middleton, Pres., Lab Dir.